



This form can be completed using Adobe Acrobat.

For your convenience, record canister #'s here

SHORT TERM Radon Gas 4" Canister INSTRUCTIONS

48 - 96 Hour Exposure Period

Get Results at: https://InspectUSA.com/results (with serial # & site zip code)

Canister 1#:	_____
Canister 2#:	_____
Canister 3#:	_____

DO NOT OPEN CANISTER UNTIL YOU ARE READY TO START TEST!

Instructions are provided to you with specific steps that must be followed. Inspect USA, National Safety Products, Accustar Labs, or any of its affiliates, cannot provide any warranty remedy to you for any claims, which arise due to the failure to follow instructions.

1. Check the expiration date on each canister. Start your test before the expiration date or results are invalid.
2. When you are ready to start the test, remove the tape from around the circumference of the canister, then remove the lid. As soon as you remove the lid, the canister is "ON" and the test has begun.
Save the lid and tape, this sheet & mailing envelope (if supplied) for returning to lab.
3. Write each canister # along with your name, address, email address, test site address & zip code on the INFORMATION FORM below. Write in the test BEGINNING date & time! Also indicate the location, floor level & the name of room (ie. basement, living room, bedroom etc) where the canister is being exposed. Be sure to print legibly.
4. PLACE THE RADON CANISTER. Canister should be placed in the lowest level of the house that is regularly used for 8-10 hours per week. If making a follow-up measurement, the US EPA recommends placing a canister on each level that is used for living space. Do NOT test in garage, porch, kitchen, closet, bathroom, furnace room, laundry room, root cellar, crawl space or sump. DO NOT place canisters where they will be exposed to high humidity &/or noticeable drafts from open doors, windows, fireplace, heat/air conditioning vents etc. Place each canister at least 3' from exterior doors or windows & at least 2' off the floor & ceiling. If performing a duplicate test (typically for Real Estate Transactions), place 2 canisters side by side 4" apart. Leave canisters in place & undisturbed for 48-96 hours.
5. END RADON TEST. After 48 hours (and before 96 hours) securely replace the lid back on each canister & replace the vinyl tape where the base meets the lid. Do not cover canister serial number with tape. Write the test ENDING date on the INFORMATION FORM below. Make sure the INFORMATION FORM is complete and LEGIBLE. If canister(s) are exposed for less than 2 days or more than 4 days the results will be invalid. Replacement canister(s) would have to be purchased & repeat the test.
6. Record the canister number(s) from the top of the can, for your reference and ability to retrieve results online.
7. Place the canister(s) & information form in the mailing package. Write your return address & seal the mailing package closed. Affix proper postage! Return IMMEDIATELY to: RADON LAB, 2 Saber Way, Ward Hill, MA 01835 DELIVERY CONFIRMATION is recommended.

Canisters must be returned within 8 days of ending the test.

If delivery of the canister is delayed more than 8 days or lost, we will not be responsible for invalid results or for a free replacement.

Reports are available same business day (approx 4.30 eastern) when canister(s) received prior to 11 AM Eastern. Canisters received after 11 AM will be available online the next business day (by 4.30 eastern).

email: results@InspectUSA.com

Get Results: https://InspectUSA.com/results (with serial # & site zip code)



CUT HERE

INFORMATION FORM

CUT HERE

Email Report To: (PRINT)		Test site address: (zip code needed for online results)	
Company:		Name:	
Name:		Address:	
Address:		City, State, Zip	
City, State, Zip		<input type="checkbox"/> Check here if canisters were placed 4" apart	
eMail address:		<input type="checkbox"/> Check here if this test is a Post Mitigation test	
Tech name & Certification # (if required):			
Canister #:	Canister #:	Canister #:	
Floor level:	Floor level:	Floor level:	
Name of area:	Name of area:	Name of area:	
Date & Time Opened:	Date & Time Opened:	Date & Time Opened:	
Date & Time Closed:	Date & Time Closed:	Date & Time Closed:	

Lab must receive within 8 days from closing the canister(s) or test is invalid.

Don't forget Postage!